

## TRIM-TO-FIT LEGPIECE AND FOOTPIECE ORDER FORM

FAX ORDER TO CUSTOMER SERVICE: 1-877-978-9703 OR EMAIL: CA.CUSTOMERSERVICE@ESSITY.COM

Patient Name: \_\_\_\_\_ DOB: \_\_\_\_\_ Date: \_\_\_\_\_

Address: \_\_\_\_\_ Gender M  F


Postal Code: \_\_\_\_\_

Diagnosis: \_\_\_\_\_

Doctor / Address: \_\_\_\_\_

Postal Code: \_\_\_\_\_

<b>PO#</b>	
Original Order <input type="checkbox"/>	Reorder w/ Changes <input type="checkbox"/>
Exact Reorder <input type="checkbox"/>	



D \_\_\_\_\_ cm C \_\_\_\_\_ cm B' \_\_\_\_\_ cm

JOBST® FARROWWRAP® TRIM-TO-FIT LEGPIECE			
Code	Size	Description	Quantity
7612300	20-60 cm (D) 20-55 (C) 20-45 (B1) 20-40 (B) length 27-35 (B-D)	LITE TTF LEGPIECE: M	
7612302	20-60 cm (D) 20-55 (C) 20-45 (B1) 20-40 (B) length 27-35 (B-D)	LITE TTF LEGPIECE EXTRA BAND: M	
7611800	20-60 cm (D) 20-55 (C) 20-45 (B1) 20-40 (B) length 27-35 (B-D)	STRONG TTF LEGPIECE: M	
7611801	45-90 cm (D) 45-90 (C) 35-80 (B1) 35-70 (B) length 27-35 (B-D)	STRONG TTF LEGPIECE: XL	
7611803	20-60 cm	STRONG TTF LEGPIECE EXTRA BAND: M	
7611804	45-90 cm	STRONG TTF LEGPIECE EXTRA BAND: XL	

JOBST® FARROWWRAP® TRIM-TO-FIT FOOT			
Code	Size	Description	Quantity
7612100	20-30 cm (A1) 14-21 (X)	LITE TTF FOOTPIECE: M	
7611600	20-30 cm (A1) 14-21 (X)	STRONG TTF FOOTPIECE: M	


Please see the back for instructions on how to trim your garment.

# Trimming and Fitting Instructions

You may visit [www.jobstcanada.com](http://www.jobstcanada.com) to view Trim-to-Fit demonstration videos.

## Trimming Instructions for LITE and STRONG Legpieces

1. Trim the distal (bottom) most band of the TTF legpiece on each side in accordance with the B circumference. Distances between lines are representative of 5 cm. If the circumference is between lines, then estimate the distance and trim accordingly.
2. Trim the bands consecutively up the legpiece corresponding to their respective L2 - L5 circumferences distance.
3. The separate band can be attached, using the Velcro® tab, to the top of the spine (in back) and moved up or down the band below as needed to increase or decrease the height of the legpiece.
4. Attach the Velcro® pieces to the patient's preferred side with exactly half extending over the edge of the band.
5. Roll the loose end of the Velcro® back onto the same band.
6. Instruct the patient not to lose the Velcro®. Replacement Velcro® sets are available for sale through your supplier.



## Trimming and Fitting Instructions for LITE and STRONG Footpiece

1. Trim the length of the footpiece to the Straight Distance X on the garment. The first line along the length represents 14 cm, and each consecutive line thereafter represents 1 additional cm.
2. Using the two numbers on either side of the footpiece closest to the toes, count out from that number and trim the width closest to the circumference at X on both sides of the footpiece on or parallel to the lines on each side of the footpiece. The distance between lines represents 2 cm. If the first line were 19, the next would be 21 and so forth. If only 1 cm increment is needed, cut in between and parallel to the lines.
3. Cut the long strip of Velcro® the same length as the straight edge you cut in step 2. Now, fold it in half and cut it in half.
4. To create the flaps for the Velcro®, fold the straight edge of the footpiece, which you just cut, on either side in half with lines facing up. Cut perpendicularly to the edge along the folded crease of the garment 2 cm deep or 3 lines. Do this on both sides.
5. Attach the Velcro® to the flaps created on patient's preferred side with half extending past the flap edge. Then fold the Velcro® back over itself onto the garment in preparation for putting it on. Make sure the patient understands that it is important they do the same.
6. For LITE TTF footpieces, trim ankle bands long enough so that the Velcro® can attach to the opposite side of the foot without touching any underlying Velcro®. Attach the included Velcro® pieces to both sides of the ankle band with half extending past the edge of the band. Then apply.
7. For STRONG TTF footpieces, to apply additional compression to the foot leave the ankle bands long. The first ankle band should be wrapped across the top of the foot to the opposite side closest to the toes, and can then be wrapped around that portion of the foot again. Affix with the Velcro®. The second ankle band should also be wrapped across to the opposite side of the foot, but closer to the ankle for additional compression to that part of the foot, or it could be used to wrap around the ankle at the discretion of the therapist. If no additional compression is desirable, they can be trimmed and applied as in step 6.

## Important Information

LITE and STRONG garments should be applied distal to proximal with ¼ overlap over the opposing bands (not including Velcro®). These garments will allow for about a 20% decrease in edema; however, keep in mind that the stated compression ranges for LITE and STRONG will be less accurate when applied to a limb larger than that for which the garment was trimmed.

- It is contraindicated to use TTF legpieces on leg circumferences less than 20 cm.
- Do not mix with non-JOBST® FarrowWrap® garments as this may increase risk to the patient and is therefore contraindicated.